

CALIBRATION CERTIFICATE

Calibration Date: December 28, 2007

Temperature: 72° F

Model Number: DS2-11

Humidity: 27% RH

Capacity: 11.00 lbf

Accuracy: 0.2% F.S.±1LSD

Resolution: 0.01 lbf

NIST Traceability

Serial Number: 229423E

Weights/Hangers #935700, #1188289, #1259740

Procedure: Imada Procedure EXA

#901384, #a1188285, #1298020

#1078843, #1297248

Temp/Humidity #15676-123519

This document certifies that the above instrument has been calibrated and tested in accordance with factory calibration procedures conducted under the conditions noted using standards which are traceable to the **National Institute of Standards and Technology (NIST)**. This document not to be reproduced without written authorization from IMADA INC

Report#122807-229423

TEST RESULTS

AS FOUND

TEST POINT (lbf)	ACTUAL (lbf)
1)	
2)	
3)	

AS LEFT

TEST POINT (lbf)	ACTUAL (lbf)
1) -11.00	-11.01
2) 0.00	0.00
3) 11.00	11.01

Signed by: _____